# PRODUCT DESCRIPTION

This low cost compact LCD voltmeter module is ideally suited for high volume applications. It features an exceptionally large display in a miniature package and includes auto-zero and user-selectable decimal points. The module is supplied with a panel mounting bezel.

#### **FEATURES**

- Low Cost
- 12.5mm (0.5") Digit Height
- 200mV d.c. Full Scale Reading
- 1mA (typ) @ +9V d.c. Power Supply
- Auto-zero and Auto-polarity
- User Selectable Decimal Points
- Supplied with Mounting Bezel



#### TYPICAL APPLICATIONS

- High Volume, Low Cost Systems
- Power Supply Monitoring
- Panel-Mount Indication

### ORDERING INFORMATION

	Stock Number
Standard Meter	V 125

### **ELECTRICAL SPECIFICATIONS**


Specification		Min.	Тур.	Max.	Unit
Accuracy (overall error) *			0.25		% (±3 counts)
Linearity	0-1000 reading		±3		counts
	0-2000 reading		±5		counts
Sample rate			3		samples/sec
Operating temperature range		0		50	°C
Temperature stability			300		ppm/°C
Supply voltage		7.0	9	12	V d.c.
Supply current			1	2	mA


<sup>\*</sup> To ensure maximum accuracy, re-calibrate periodically.


### **SAFETY**


To comply with the Low Voltage Directive (LVD 93/68/EEC), input voltages to the module's pins must not exceed 60Vdc. The user must ensure that the incorporation of the panel meter into the user's equipment conforms to the relevant sections of BS EN 61010 (Safety Requirements for Electrical Equipment for Measuring, Control and Laboratory Use).














# **PIN FUNCTIONS**

1. V+ Positive power supply connection.

2. V- Negative power supply connection.

3. INHI Positive measuring input with reference to INLO. Analogue inputs must be no closer than 1V

4. INLO Negative measuring input with reference to IN HI. to either the positive or negative supply.
5. XDP Connect to DP1, 2 or 3 to display required decimal point.

6. DP1 Connect to XDP to display Decimal Point 1 (199.9).

7. DP2 Connect to XDP to display Decimal Point 2 (19.99).

8. DP3 Connect to XDP to display Decimal Point 3 (1.999).

# Solder Links:

L1 Normally Closed, connects COM to INLO.




### **SCALING**

Two resistors Ra and Rb may be used to alter the full scale reading (FSR) of the meter - see table. The meter will have to be recalibrated by adjusting the calibration potentiometer on the rear of the module.

	FSR	Ra	Rb
Voltage Vin	2V	910k	100k
	20V	1M	10k
	200V	1M	1k
	2000V*	1M	100R
Current lin	200μΑ	0R	1k
	2mA	0R	100R
	20mA	0R	10R
	200mA	0R	1R







# **APPLICATIONS**

Do not connect more than one meter to the same power supply if the meters cannot use the same signal ground. Taking any input beyond the power supply rails will damage the meter.






Measuring an input voltage referenced to a floating supply, i.e. the input voltage and the meter's power supply are isolated from each other. Check Link L1 is closed.



Measuring a current referenced to a floating supply, i.e. the current being measured and the meter's power supply are isolated from each other. Check Link L1 is closed.



Measuring a single ended input referenced to supply, i.e. the input voltage and the meter's power supply share the same 0V rail. Check Link L1 is open.

Specifications liable to change without prior warning

V 125

Issue 3

January/2002

M.C.

Applies to V 125/2

